

Determination of titanium content in pyrotechnics used for fireworks and firecrackers based on Energy Dispersive X-ray Fluorescence Spectrometry (EDXRF)

Wu Jun-yi^{1, a}, XiaoHuan-xin^{2, b}

¹Technical Center for Dangerous Goods Testing of Guangxi Entry-Exit Inspection and Quarantine Bureau, Beihai, Guangxi, China

^ajohnny-woo@vip.163.com

^bxhxciq@china.com

Keywords: fireworks and firecrackers; pyrotechnics; titanium content; determination; Energy Dispersive X-ray Fluorescence Spectrometry; EDXRF.

Abstract. Methods used for the determination of titanium content in pyrotechnics are mostly based on traditional chemical method, which is lengthy and cumbersome. If inductively coupled plasma emission spectrometry or atomic absorption spectrometry are used to determine the titanium with high content, the sample solution must be highly diluted, and it must produce errors in measurement and calling into question the reliability of the data. The method mentioned in this paper is about the determination of titanium content in pyrotechnics used for fireworks and firecrackers based on energy dispersive X-ray fluorescence spectrometry by controlling matrix effects between elements. Using sample solution of pyrotechnics in specific concentrates, the titanium content can be determined by the specific calibration curve established with an intensity calibration. This method can provide high accuracy and good precision in a short time with a simple process by efficiently controlling the matrix effects. It can fully meet the requirements for the determination of titanium in pyrotechnics used for different kinds of fireworks and firecrackers around the world, and it has good generalization and practicability. The average recovery of the method can be 98.88% ~ 100.56%, allowing for a difference of 0.5%.

Introduction

In China, fireworks and firecrackers are very important consumer recreational products in people's everyday life since ancient times. Gorgeous colors produced by fireworks and firecrackers are even the titaniuming role of foiling festal atmosphere in every grand holiday celebrations. In recent years, with the rapid development of global trade, fireworks and firecrackers are becoming more and more popular all over the world, more and more consumers are fascinated by different kinds of patterns, pictures, and sound effects of fireworks and firecrackers. Titanium element is commonly found in titanium powder as primary content used for pyrotechnics. Quantitative analysis of chemical compositions in pyrotechnics such as titanium content is required under the Globally Harmonized System of Classification and Labeling of Chemicals (GHS) to be complemented in the fireworks and firecrackers industry. Meanwhile it will also provide a scientific and effective technical support to the management and supervision of safety production for the government, and improve products' quality level by the manufacturers. It can also be utilized as a tool in providing valuable data in the judgment in some major arbitration and security incident analysis. Quantitative analysis method of the titanium content reported in current literature is limited to traditional chemical analysis, such methods have the following disadvantages: (1) Long detecting period. Generally, it will take a skilled technician two whole days or so to complete the detection. (2) The operation is more complicated. It needs to go through many steps such as dissolving sample, filtration, precipitation collection, drying and weighing precipitation and ect. Comparing with traditional chemical analysis methods, this method based on energy dispersive X-ray fluorescence spectrometry (EDXRF) has the advantages of simple operation steps, short period of detection, high accuracy and good precision.

Theory

Titanium element is commonly found as primary content in chemical materials such as titanium powder in pyrotechnics used for fireworks and firecrackers. Statistical analysis shows that titanium powder in pyrotechnics is between 5% to 60% ,it can concludes that the titanium content in pyrotechnics would be 2.5% ~60% as mass fraction. Concept of the method: considering the weight of the sample is 1.5g, constant volume is 0.5L and the concentrations of the titanium would be controlled in 0.25g/L ~0.60g/L in sample solutions. And it can prove that when the titanium content in the solution is in the range of 0.12 g/L ~1.12g/L, there would be little matrix effects among elements. So we can establish a working curve which contains the titanium elements with the content of 0.12 g/L ~1.12g/L to determine the titanium content in the sample solution. In accordance with the relevant safety regulations, the sample was ground into powder of less than 180 micron. Then the sample powder is placed in an explosive-proof oven at 50°C–55°C and dried for 4 hours, and then placed into a dryer for cooling down to room temperature. Pretreated sample is fully dissolved in 150mL pure water and then filtered into volumetric flask as sample solution. The sample solution can be put into the sample cup and placed in the tank of the EDXRF to measure the fluorescence intensity of the titanium elements. The actual content of titanium element in the sample can be calculated from the concentrations of the titanium reading by the working curve.

Experiment section

Reagents

Unless otherwise stated, all the reagents should be guaranteed reagents and pure water is secondary grade water as described in ISO 3696(1987). Hydrofluoric acid. Standard working solution of the titanium fluoride: Weigh 1.26 g high purity titanium powder reference materials (accuracy to 0.1 mg), and put it in a 300 ml beaker, add 50mL pure water and 35 mL hydrofluoric acid, heat the beaker and make the sample solution slightly boiling on an electric stove for 30 min. After the solution is cool down to the room temperature, transfer the solution into a 500 ml volumetric flask and add pure water to the scale. Then we can separately pipette the standard working solution of the titanium fluoride with volume 5mL、10mL、15mL、20mL、25mL、30mL、35mL、40mL and 45mL into nine 100mL volumetric flasks which marked from N₁ to N₉, and add pure water to reach 100 mL in each volumetric flask, mix thoroughly for later use. Concentrations of the standard working solution in different flasks are shown in Table 1.

Table 1 Concentrations of series standard working solutions of titanium fluoride

NO.	Mass concentration (g/L)	Fluorescence intensity (cps/mA)
N ₁	0.1246	62.83
N ₂	0.2492	129.06
N ₃	0.3739	193.26
N ₄	0.4985	261.77
N ₅	0.6231	315.57
N ₆	0.7477	382.70
N ₇	0.8723	445.74
N ₈	0.9970	514.63
N ₉	1.1216	579.02

Instrument and apparatus

Explosive-proof oven with accuracy to $\pm 2^{\circ}\text{C}$. Analytical balance with accuracy to 0.1 mg. energy dispersive X-ray fluorescence spectrometer (EDXRF): United States Thermo Fisher (former Thermo Electron Corporation) Company QUANT'X series.

Operation step

- (1) Weigh the sample of about 2.0 g, accuracy to 0.1 mg.
- (2) Place the sample into a clean 300 mL beaker, add 150mL pure water into the beaker, then place the beaker on an electric stove to make the solution boiled for about 10 min. Filter the solution through filter paper and collect the residue to another 300 mL beaker, add 60mL hydrofluoric acid and 40 mL pure water to the beaker, then place the beaker on an electric stove to make the solution boiled for about 20 min. Filter the solution through filter paper to an 0.5L volumetric flask, wash the beaker and the filter paper several times with pure water, and make a constant volume with pure water after the filtered solution cooling down to room temperature.
- (3) Parameters of the EDXRF instrument parameters are shown in Table 2.

Table 2Parameters of the EDXRF instrument

Filter	Aluminum
Collimator	8.8mm
Voltage	12v
Electric current	Auto
Analysistime	50s
Count rate	Medium
Atmosphere	Air
Matrix effects	Not considered
Energy range	0~20kev
Analysis technique	Intensity correction
sample thickness	$\geq 15\text{mm}$

- (4) Calibration (working) curve: according to the requirements of the method and the instrument criteria, we set the instrument to optimum analysis conditions, and adjust it to the best working condition, and determine spectral intensity of the series standard solution from N_1 to N_9 to establish the calibration (working) curve with the elemental concentrations as independent variable and the spectral intensity as the dependent variable. The linear correlation coefficient of the regression curve should be 0.99 or higher.
- (5) Sample determination: determine the fluorescence intensity of the titanium in blank solution and every sample solution under the best analysis condition and read the concentrations from the calibration curve according to the spectral intensity.

Results calculation

Content of the titanium element in the sample can be calculated as mass fraction W and its value shown in% according to the following formula.

$$\omega = \omega_0 \times \frac{2}{m} \times \frac{V}{500}$$

Where: ω_0 —the content of the titanium in the sample read by the working curve, expressed in %.
 m —quantity of the sample, expressed in milligrams (g).

V —constant volume of the volumetric flask used for the sample solution, expressed in liters(mL).

ω — the content of the titanium in the sample, expressed in %.

z —assume that quantity of the sample, expressed in milligrams (g).

500—assume that constant volume of the volumetric flask used for the sample solution, expressed in liters(mL).

Results and discussion

Solvent selection

Considering the characteristics of the EDXRF spectrometry, this method selects hydrofluoric acid as the solvent for the sample instead of some other strong acids such as hydrochloric acid, sulfuric acid, or perchloric acid, which are usually recommended in relevant papers. If these strong acids were to be selected as the solvents to dissolve the sample, great amounts of chlorine and sulfur elements would be introduced to the sample solution, and these would make great matrix effects on the titanium element and affect the accuracy of the test. On the contrary, if hydrofluoric acid are used as the solvents, only the fluorine elements are introduced to the sample solution. So, the other elements would have little matrix effects on the titanium element and can be basically ignored.

Selection of standard solution.

Considering that all the most of the titanium element come from titanium powder in pyrotechnics used for fireworks and firecrackers. In order to make the standard solution as consistent as possible with the sample solution, the titanium standard solution would be selected to make the working curve. It proved that when the concentration of the titanium element is controlled to the range of 0.12 g/L ~ 1.12 g/L, it would have little matrix effects on the titanium element and can be basically ignored. Because the contents of other impurity elements such as aluminum and manganese are all mostly less than the titanium element, they would also have little matrix effects on the titanium element in the sample solution.

Recovery test

To assess the accuracy of the method, we used the standard reference substances addition recovery test. We added the titanium reference standard substance to some actual black powder samples, dissolved the samples and determined the contents of titanium in the sample solution. The values of the titanium contents we measured are compared with the theoretical ones, and the experimental data is shown in Table 3 below.

Table 3 Recovery test results

NO.	Reference materials	Reference Code	Titanium content of nominal (%)	Titanium content of measurement (%)	recovery rate (%)
1	Black powder	T01	12.38	12.25	98.95
2	Lifting charge	T02	16.22	16.12	99.38
3	Bursting charge	T03	25.28	25.16	99.53
4	Titanium standard solution	T04	7.12	7.06	99.16
5	Titanium standard solution	T05	11.46	11.49	100.26
6	Titanium standard solution	T06	15.25	15.37	100.79
7	Titanium standard solution	T07	21.38	21.23	99.30
8	Titanium standard solution	T08	24.86	24.75	99.56
9	Titanium standard solution	T09	27.54	27.42	99.56
10	Titanium standard solution	T10	28.12	28.01	99.61
Average (X)					99.61
standard deviation (S)					0.51

Conclusions

Energy dispersive X-ray fluorescence spectrometry (EDXRF) is used to determine the titanium content in pyrotechnics used for fireworks and firecrackers, this method is accurate and quick with high accuracy and good precision. When the titanium content in the sample is in the range of 7% ~ 28% as mass fraction, the recovery is 98.95% ~ 100.79%. The allowable differential value was 0.5% between two single tests under repeatable conditions. In other words, this method can completely satisfy the requirements of the fireworks and firecrackers industry.

References

- [1] Leif, H. C.; Allan, A. Determination of sulfur and heavy metals in crude oil and petroleum products by energy-dispersive x-ray fluorescence spectrometry and fundamental parameter approach. *Anal. Chem.*, 53(12), pp.1788-1792, 1981.
- [2] Fei, H; Pierre, J. V. E. General approach for quantitative energy dispersive x-ray fluorescence analysis based on fundamental parameters. *Anal. Chem.*, 63(20), pp.2237-2244, 1991.
- [3] Standard test method for determination of lead in paint layers and similar coatings or in substrates and homogenous materials by energy dispersive x-ray fluorescence spectrometry using multiple monochromatic excitation beams, ASTM F2853, American Society for Testing and Materials Publications, 2010.
- [4] Standard test method for analysis of uranium and thorium in soils by energy dispersive x-ray fluorescence spectroscopy, ASTM C1255, American Society for Testing and Materials Publications, 2011.
- [5] Duan Tiyu, Li Suqing, et al. Jewellery, Determination of precious metal content - Method using x-ray fluorescence spectrometry, China National Recommended Standard: GB/T 18043-2013, China Standards Press: Beijing, 2014.

- [6] Leoni. L,Saitta. M, X-ray fluorescence analysis of powder pellets utilizing a small quantity of material, X-ray Spectrom, 3,pp.74-77,1974.
- [7] Rose W.I.,Bornhorst T.J.,Sivonen S.J., Rapid high-quality major and trace element analysis of powdered rock by x-ray fluorescence spectrometry, X-ray Spectrom, 15,pp.55-60,1986.
- [8] Gy,PierreM,The analytical and economic importance of correctness in sampling,Anal.Chim.Acta, 190,pp.13-23,1986.
- [9] ZHOU Tong-hui ,WANG Er-kang,LU Wan-zhen ect. Handbook in analytical chemistry(second edition), basic knowledge of and safety knowledge(the first volume) , China Chemical industry press:pp.568-580,1997.